

Betriebsdaten:
Typical operation:
Exemple d'utilisation:

12	A 109
U_f . . . 1,1	. . . 1 V
I_f . . . 0,25	. . . 0,5 A
U_a . . +135	. +120 V=
U_g . . -10,5	. -0 V=
I_a . . . 3	. . . 5 mA
S . . . 0,44	. . . 0,5 mA/V
D 15	. . . 10 %
g 6,6	. . . 10
R_i . . . 15 k Ω

Prüfdaten:

Test operation:

Vérification:

U_f . . .	1 V~
U_a . . .	+150 V=
U_g . . .	0 & -2 V=

In Stellung 13
auf Steuerwirkung
prüfen

In position 13 is to
prove the control grid.
En position 13 vérifier
l'action de commande de la
grille.

612

Patent-Röhrenprüfer Modell W 19

MAX FUNKE, Adenau-Eifel 11. 54.


12	1120	FP 56		A 109	H 107	
112	C 12	VT-7	WX 12	C 109	H 125	REN 511
412	F	WD 12	264	G 115	R 150	

Triode.

Unbrauchbar
mauvais / Bad

?

G u t
Good / Bon



10 mA

