

**Betriebsdaten:**  
 Typical operation:  
 Exemple d'utilisation:

$U_f$  . . . 6,3 V  
 $I_f$  . . . 0,3 A  
 $U_a$  . max 150 V  
 $I_a$  . max 0,6 mA  
 $U_{fk}$  . max 90 V

**Prüfdaten:**  
 Test operation:  
 Vérification:

$U_f$  . . . 6,3 V~  
 $U_a$  . . . 10 V~

**755**



Röhre hat **2 Systeme**.  
 Das 2. System ist in  
 Stellung **11** zu messen  
 2. System is measurable  
 in position **11**  
 2. System est à essayer  
 en position **11**

Patent-Röhrenprüfer Modell W19

MAX FUNKE, Adenau-Eifel 2. 54.

**6 SQ 7**

6 SQ 7 G    6 SQ 7 GT/G  
 6 SQ 7 GT    VT-103

Triode + Duodiode  
 Karte 754 + Karte 755

**6 SR 7**

6 SR 7 G  
 6 SR 7 GT  
 VT-233

**6 ST 7**



6 ST 7 G  
 VT-205

Duodiode + Triode  
 Karte 755 + Karte 756

Unbrauchbar  
 mauvais / Bad

?

**G u t**  
 Good / Bon



**1 mA**