



12 SA 7

Heptode

12 SA 7 G VT 161

Hauptinstrument
Bereich



12 SA 7 GT 12 SY 7

12 SY 7 GT

● **6mA**

● A 250 Volt

● U_{II} 0 Volt

Sollwert: 3,4mA

● G_3

● U_{III} 100 Volt

● G_{2+4}

● U_I 2 Volt

● G_1

● M, (G_5)

● K, (G_5)

● F

Heizung: 12,6 Volt

● F

Prüfkarte für RPM 370
Neuberger, München 25